

UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE United States Patent and Trademark Office Address: COMMISSIONER FOR PATENTS P.O. Box 1450 Alexandria, Virginia 22313-1450 www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/076,369	02/19/2002	Henry William Lupton	Q68587	1502
7590 11/25/2005			EXAMINER	
SUGHRUE MION, PLLC 2100 Pennsylvania Avenue, NW Washington, DC 20037-3213			HO, UYEN T	
			ART UNIT	PAPER NUMBER
			3731	
			DATE MAIL FD: 11/25/2005	


Please find below and/or attached an Office communication concerning this application or proceeding.

		Applicant(s)			
Office Action Summan	10/076,369	LUPTON, HENRY WILLIAM			
Office Action Summary	Examiner	Art Unit			
	(Jackie) Tan-Uyen T. Ho	3731			
The MAILING DATE of this communication app Period for Reply	pears on the cover sheet with the c	orrespondence address			
A SHORTENED STATUTORY PERIOD FOR REPL WHICHEVER IS LONGER, FROM THE MAILING D - Extensions of time may be available under the provisions of 37 CFR 1.1 after SIX (6) MONTHS from the mailing date of this communication. - If NO period for reply is specified above, the maximum statutory period of Failure to reply within the set or extended period for reply will, by statute Any reply received by the Office later than three months after the mailine earned patent term adjustment. See 37 CFR 1.704(b).	ATE OF THIS COMMUNICATION 36(a). In no event, however, may a reply be tin will apply and will expire SIX (6) MONTHS from , cause the application to become ABANDONE	N. nely filed the mailing date of this communication. D (35 U.S.C. § 133).			
Status					
1) Responsive to communication(s) filed on <u>09 September 2005</u> . 2a) This action is FINAL . 2b) This action is non-final. 3) Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under <i>Ex parte Quayle</i> , 1935 C.D. 11, 453 O.G. 213.					
Disposition of Claims					
4) ⊠ Claim(s) 1-6 is/are pending in the application. 4a) Of the above claim(s) 3-5 is/are withdrawn from consideration. 5) □ Claim(s) is/are allowed. 6) ⊠ Claim(s) 1.2.6 is/are rejected. 7) □ Claim(s) is/are objected to. 8) □ Claim(s) are subject to restriction and/or election requirement.					
Application Papers					
9) The specification is objected to by the Examine 10) The drawing(s) filed on is/are: a) acc Applicant may not request that any objection to the Replacement drawing sheet(s) including the correct 11) The oath or declaration is objected to by the E	epted or b) objected to by the drawing(s) be held in abeyance. Se tion is required if the drawing(s) is ob	e 37 CFR 1.85(a). ojected to. See 37 CFR 1.121(d).			
Priority under 35 U.S.C. § 119					
 12) Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f). a) All b) Some * c) None of: 1. Certified copies of the priority documents have been received. 2. Certified copies of the priority documents have been received in Application No. 3. Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)). * See the attached detailed Office action for a list of the certified copies not received. 					
Attachment(s) 1) Notice of References Cited (PTO-892) 2) Notice of Draftsperson's Patent Drawing Review (PTO-948) 3) Information Disclosure Statement(s) (PTO-1449 or PTO/SB/08 Paper No(s)/Mail Date	6) Other:				

DETAILED ACTION

Response to Arguments

1. Applicant's arguments filed 9/9/05 have been fully considered but they are not persuasive. Applicants argue that Dusbabek et al. does not teach or suggest a catheter and stent combination. Examiner disagrees. Figures 1-8, 22-56, Dusbabek et al. disclose catheter and stent combination. Applicant also argues that Dusbabek et al. does not disclose the configuration of the catheter that meets the term corrugated as being "to shape into folds or parallel and alternating ridges and grooves." Examiner disagrees. Figures 45 of Dusbabek et al. reference as illustrated below



These ridges and grooves are parallel.

2. The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that

form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless -

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United

States.

3. Claims 1, 2, 6 are rejected under 35 U.S.C. 102(b) as being anticipated by

Dusbabek et al. (5,968,069). Regarding to claims 1, 2 and 6, Dusbabek et al. disclose

a catheter having an elongate body (212), the body (212) comprising a tubular member

(216) surrounded the elongate body wherein the walls of the tubular member is

corrugated and including a plurality of ribs (fig. 30-33), a stent (218) and a support

region (214) is formed over the corrugated section.

Regarding to claims 1 and 6, Dusbabek et al. also discloses a catheter having an

elongate body (372, fig. 43-45), the body being a tubular member and the tubular

member having at least one corrugated section (354) in the distal region, a support

region (358) formed over the corrugated section and a stent (360).

Conclusion

4. THIS ACTION IS MADE FINAL. Applicant is reminded of the extension of time

policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire THREE

MONTHS from the mailing date of this action. In the event a first reply is filed within

TWO MONTHS of the mailing date of this final action and the advisory action is not

mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the mailing date of this final action.

5. Any inquiry concerning this communication or earlier communications from the examiner should be directed to (Jackie) Tan-Uyen T. Ho whose telephone number is 571-272-4696. The examiner can normally be reached on MULTIFLEX Mon. to Sat.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, ANHTUAN NGUYEN can be reached on 571-272-4963. The fax phone number for the organization where this application or proceeding is assigned is 703-872-9306.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

(Jackie) Tan-Uyen T. Ho Patent Examiner

Art Unit 3731